

Strand-Specific Reverse Transcription PCR for Detection of Replicating SARS-CoV-2

Catherine A. Hogan, ChunHong Huang, Malaya K. Sahoo, Hannah Wang, Becky Jiang, Mamdouh Sibai, Marisa Holubar, Roshni Mathew, James Zehnder, Benjamin A. Pinsky

We developed an assay that detects minus-strand RNA as a surrogate for actively replicating severe acute respiratory syndrome coronavirus 2. We detected minus-strand RNA in 41 persons with coronavirus disease up to 30 days after symptom onset. This assay might inform clinical decision-making about patient infectiousness.

Real-time reverse transcription PCR (rRT-PCR) is the standard diagnostic method for coronavirus disease 2019, but it cannot differentiate between actively replicating and inactive virus. Active replication is a critical factor for infectiousness; however, its time course is difficult to estimate because of the typical 20–50 days before rRT-PCR negative conversion occurs (1,2). PCR cycle threshold (C_t) values might help physicians to determine a patient's infectiousness, but researchers have isolated replicating virus from patients with a wide range (28–33) of C_t values (3–7). Given the stringent biosafety precautions needed for viral culturing of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), physicians require additional diagnostic tools. Actively replicating virus produces minus-strand RNA intermediates that can be detected by PCR (8,9). We developed and validated a 2-step strand-specific rRT-PCR for the detection of actively replicating SARS-CoV-2 and assessed its clinical performance.

The Study

We conducted standard nucleic acid and amplification testing at the Stanford Health Care Clinical Virology Laboratory (Stanford, CA, USA) using the Panther Fusion SARS-CoV-2 Assay (Hologic Inc., <https://www.hologic.com>), the Panther Aptima SARS-CoV-2 Assay (Hologic Inc.), or the in-house

rRT-PCR specific to the SARS-CoV-2 envelope gene (permitted by Emergency Use Authorization) (10,11). We did not culture SARS-CoV-2 because we did not have access to a biosafety level 3 laboratory.

We developed a novel 2-step rRT-PCR specific to the minus strand of the envelope gene (Appendix, <https://wwwnc.cdc.gov/EID/article/27/2/20-4168-App1.pdf>). First, we used strand-specific primers to convert SARS-CoV-2 RNA to complementary DNA. Then, we amplified the complementary DNA by rRT-PCR in 3 separate positive, negative, and background (no primer) reactions using the Rotor-Gene Q instrument (QIAGEN, <https://www.qiagen.com>) (Appendix). We conducted the analytical validation during May–June 2020. We used in vitro transcribed minus- and plus-strand RNA to evaluate the linearity, precision, and lower limit of detection of the assay (Appendix).

We retrospectively collected a convenience set of upper respiratory specimens with a broad range of C_t values. These samples had been collected and frozen from 93 inpatients and outpatients who were treated at Stanford Health Care and tested positive for SARS-CoV-2 during March 12–April 9, 2020. We also reviewed the electronic medical records of the participating patients. For the prospective phase of the study, we collected upper respiratory samples from 53 consecutive patients with confirmed SARS-CoV-2 infection by standard rRT-PCR during July 31–September 4, 2020 (Appendix). Treating physicians ordered strand-specific rRT-PCR on the basis of clinical need; we used samples from these patients in the prospective phase.

We conducted analytical validation (12) and statistical analysis using Stata version 15.1 (StataCorp LLC., <https://www.stata.com>) (Appendix). We considered a 2-tailed $p < 0.05$ to be significant. This study was approved by the Stanford Institutional Review Board (protocol no. 48973).

In total, we analyzed specimens from 146 patients: 93 in the retrospective phase and 53 in the prospective

Author affiliations: Stanford University School of Medicine, Stanford, California, USA (C.A. Hogan, C. Huang, M.K. Sahoo, H. Wang, M. Holubar, R. Mathew, J. Zehnder, B.A. Pinsky); Stanford Health Care, Stanford (C.A. Hogan, B. Jiang, M. Sibai, B.A. Pinsky)

DOI: <https://doi.org/10.3201/eid2702.204168>




Figure 1. Frequency distribution of days between symptom onset and testing in study on strand-specific real-time reverse transcription PCR for detection of replicating severe acute respiratory syndrome coronavirus 2, California, USA, 2020. Dashed line indicates the median number of days since symptom onset. B) Distribution of standard real-time reverse transcription PCR cycle threshold values by results of strand-specific real-time reverse transcription PCR. Horizontal line indicates median.

phase (Appendix Tables 3, 4). The median age was 50 years (interquartile range 36–63 years); 73 (50.0%) were women, 26 (17.8%) were immunocompromised, and 30 (20.5%) were admitted to the intensive care unit for coronavirus disease during the course of the study (Table 1). Samples were collected a median of 9 days (interquartile range 4–18 days) after symptom onset (Figure 1, panel A). We detected minus-strand RNA in 41 (28.1%) patients. The median C_t value of samples with detected minus-strand RNA (20.7) was significantly lower than those in which the minus strand was not detected (33.2; $p < 0.01$) (Figure 1, panel B). The results of this strand-specific assay were closely correlated with the standard rRT-PCR results (Figure 2, panels A, B). The ratio of minus:plus strands varied by patient within 14 days after symptom onset (Appendix Figure 2).

We detected the minus strand in 7 patients in the prospective cohort (Table 1, <https://wwwnc.cdc.gov/EID/article/27/2/20-4168-T1.htm>). Two of these patients were nonimmunocompromised inpatients tested >10 days after symptom onset, including 1 who had been asymptomatic for >48 hours; the C_t values for these samples were 39.0 and 38.6. We detected minus-strand SARS-CoV-2 RNA up to 30 days after symptom onset in an immunocompromised patient with

persistent fever. For 2 patients in the prospective cohort, a negative result might have facilitated the approval of medical procedures despite prolonged positive results by standard rRT-PCR (Appendix).

Conclusions

We described the performance of a 2-step strand-specific rRT-PCR for detection of SARS-CoV-2. The assay identified viral replication in patients with persistent positive results by standard rRT-PCR, possibly facilitating clinical decision-making.

Other assays that assess intermediates of viral replication, such as subgenomic RNA, have emerged in the literature (5,13). Perera et al. demonstrated high correlation between levels of presumptive SARS-CoV-2 active replication intermediates and standard rRT-PCR C_t values (13). The standard SARS-CoV-2 rRT-PCR is appropriate for most routine clinical diagnostic applications. However, because this assay does not determine whether SARS-CoV-2 is actively replicating, it cannot infer infectiousness in samples with mid-level C_t values (i.e., C_t 25–35).

We detected minus-strand RNA up to 30 days after symptom onset, which is longer than the 14-day period previously reported for subgenomic RNA (13),




Figure 2. Deming regression analysis of C_t values by strand-specific real-time reverse transcription PCR as a function of the C_t values by standard real-time reverse transcription PCR for severe acute respiratory syndrome coronavirus 2. Results of PCR for plus strand (A; $y = 0.91x + 3.26$) and minus strand (B; $y = 0.88x + 17.30$). C_t , cycle threshold.

Table 2. Clinical characteristics of 7 patients with detected SARS-CoV-2 minus-strand RNA, California, USA, 2020*

| Patient ID | Age, y/sex | Immuno-compromised | Test order | Cycle threshold value for standard rRT-PCR specific to SARS-CoV-2 | Days after symptom onset | Symptomatic improvement at time of strand-specific testing† | Fever within 24 h of strand-specific testing | Minus strand detected beyond CDC isolation recommendations‡ |
|------------|------------|--------------------|------------|---|--------------------------|---|--|---|
| 102 | 75/M | No | 1 | 33.5 | Unclear | Unclear§ | No | Unclear |
| | | | 2 | 19.7 | Unclear | Unclear§ | No | Unclear |
| 111 | 58/M | Yes | 1 | NA¶ | 26 | No | Yes | No |
| 117 | 82/M | No | 1 | 18.5 | 12 | Unclear§ | No | Yes |
| 118 | 69/M | No | 1 | 20.8 | 4 | No | Yes | No |
| 127 | 61/M | No | 1 | 34.5 | 11 | Yes | No | Yes |
| 129 | 60/M | Yes | 1 | 22.6 | 18 | No | Yes | No |
| | | | 2 | 20.2 | 30 | No | Yes | No |
| 141 | 2/F | Yes | 1 | 18.1 | NA# | NA# | NA# | NA# |

*ID, identification; NA, not available; rRT-PCR, real-time reverse transcription PCR; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2.
†From symptom onset to time of testing.
‡According to recommendations from the Centers for Disease Control and Prevention (14).
§Because of underlying condition.
¶Qualitative testing conducted at external reference laboratory.
#Data not available because patient was treated at external medical facility.

and 8–15 day period for viral culture (3–6,13). We detected minus-strand RNA in 2 patients beyond the typical period recommended for isolation. Isolation strategies on the basis of time and symptoms are simple to apply, reduce the number of tests that need to be conducted, thus saving resources, and are probably effective at a population level (14). However, it can be challenging to determine the infectiousness of patients in certain clinical contexts, such as immunocompromised hosts with persistent viral shedding, on the basis of time and symptoms alone. Tools such as strand-specific RNA testing might be helpful in determining the infectiousness of these patients. Strand-specific testing might also help avoid delays in required procedures or treatments such as chemotherapy, which might be postponed because of SARS-CoV-2–positive PCR results.

This study has several strengths, including a large patient cohort and analytical validation. This strand-specific assay is useful because it can be adapted for routine clinical laboratory testing, does not require emergency use authorization, and reports C_t values and strand-specific RNA detection. The study was limited by its single-center design and combination of 2 patient cohorts chosen using different selection techniques. The assay lacks viral culture data and is hampered by longer turnaround time and complexity. In future studies, we will validate this assay against SARS-CoV-2 viral culture and within a household transmission study.

In summary, we described the test performance and clinical feasibility of a strand-specific rRT-PCR assay for SARS-CoV-2. Strand-specific rRT-PCR testing might be especially useful in patients with prolonged RNA shedding. It might also supplement existing strategies for estimating infectiousness on the basis of time and symptoms. Further work is required to correlate these findings with viral culture, compare

different strand-specific RNA detection methods, and to assess clinical utility in large and longitudinal patient cohorts. These findings might improve understanding of the infectiousness of SARS-CoV-2, enabling optimization of infection control measures and resource use.

Acknowledgments

We thank the staff from the Stanford Clinical Virology Laboratory for their testing of samples.

About the Author

Dr. Hogan is a medical microbiologist in the Department of Pathology, Stanford University, Palo Alto, CA. Her research interests include novel and point-of-care diagnostics, clinical impact of diagnostic methods, and global health.

References

- Gombar S, Chang M, Hogan CA, Zehnder J, Boyd S, Pinsky BA, et al. Persistent detection of SARS-CoV-2 RNA in patients and healthcare workers with COVID-19. *J Clin Virol*. 2020;129:104477. <https://doi.org/10.1016/j.jcv.2020.104477>
- Sun J, Xiao J, Sun R, Tang X, Liang C, Lin H, et al. Prolonged persistence of SARS-CoV-2 RNA in body fluids. *Emerg Infect Dis*. 2020;26:1834–8. <https://doi.org/10.3201/eid2608.201097>
- Bullard J, Dust K, Funk D, Strong JE, Alexander D, Garnett L, et al. Predicting infectious severe acute respiratory syndrome coronavirus 2 from diagnostic samples. *Clin Infect Dis*. 2020;ciaa638. <https://doi.org/10.1093/cid/ciaa638>
- La Scola B, Le Bideau M, Andreani J, Hoang VT, Grimaldier C, Colson P, et al. Viral RNA load as determined by cell culture as a management tool for discharge of SARS-CoV-2 patients from infectious disease wards. *Eur J Clin Microbiol Infect Dis*. 2020;39:1059–61. <https://doi.org/10.1007/s10096-020-03913-9>
- Wölfel R, Corman VM, Guggemos W, Seilmaier M, Zange S, Müller MA, et al. Virological assessment of hospitalized patients with COVID-2019. *Nature*. 2020;581:465–9. <https://doi.org/10.1038/s41586-020-2196-x>

6. Jeong HW, Kim SM, Kim HS, Kim YI, Kim JH, Cho JY, et al. Viable SARS-CoV-2 in various specimens from COVID-19 patients. *Clin Microbiol Infect.* 2020;26:1520–4. <https://doi.org/10.1016/j.cmi.2020.07.020>
7. Tom MR, Mina MJ. To interpret the SARS-CoV-2 test, consider the cycle threshold value. *Clin Infect Dis.* 2020;71:2252–4. <https://doi.org/10.1093/cid/ciaa619>
8. Sztuba-Solińska J, Stollar V, Bujarski JJ. Subgenomic messenger RNAs: mastering regulation of (+)-strand RNA virus life cycle. *Virology.* 2011;412:245–55. <https://doi.org/10.1016/j.virol.2011.02.007>
9. Sawicki SG, Sawicki DL, Siddell SG. A contemporary view of coronavirus transcription. *J Virol.* 2007;81:20–9. <https://doi.org/10.1128/JVI.01358-06>
10. Hogan CA, Sahoo MK, Huang C, Garamani N, Stevens B, Zehnder J, et al. Comparison of the Panther Fusion and a laboratory-developed test targeting the envelope gene for detection of SARS-CoV-2. *J Clin Virol.* 2020;127:104383. <https://doi.org/10.1016/j.jcv.2020.104383>
11. U.S. Food and Drug Administration. Stanford Health Care Clinical Virology Laboratory SARS-CoV-2 test EUA Summary. 2020 Nov 24 [cited 2020 Dec 2]. <https://www.fda.gov/media/136818/download>
12. Chesher D. Evaluating assay precision. *Clin Biochem Rev.* 2008;29:S23–6.
13. Perera RAPM, Tso E, Tsang OTY, Tsang DNC, Fung K, Leung YWY, et al. SARS-CoV-2 virus culture and subgenomic RNA for respiratory specimens from patients with mild coronavirus disease. *Emerg Infect Dis.* 2020;26:2701–4. <https://doi.org/10.3201/eid2611.203219>
14. U.S. Centers for Disease Control and Prevention. Discontinuation of Transmission-Based Precautions and Disposition of Patients with COVID-19 in Healthcare Settings (Interim Guidance). 2020 [cited 2020 Sep 13]. <https://www.cdc.gov/coronavirus/2019-ncov/hcp/disposition-hospitalized-patients.html>

Address for correspondence: Benjamin A. Pinsky, Stanford University School of Medicine, 3375 Hillview Avenue, Room 2913, Palo Alto, CA 94304, USA; email: bpinsky@stanford.edu



EID
journal

@CDC_EIDJournal

Want to stay updated on the latest news in *Emerging Infectious Diseases*? Let us connect you to the world of global health. Discover groundbreaking research studies, pictures, podcasts, and more by following us on Twitter at @CDC_EIDJournal.

Strand-Specific Reverse Transcription PCR for Detection of Replicating SARS-CoV-2

Appendix

Routine SARS-CoV-2 Testing

All testing was performed at the Stanford Clinical Virology Laboratory (Stanford, CA, USA), a laboratory in northern California that serves 2 academic medical centers and affiliated clinics in the surrounding area. Standard clinical testing of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) from respiratory samples was performed by the emergency use authorization for real-time reverse transcription PCR (rRT-PCR) specific to the SARS-CoV-2 envelope (*E*) gene, or with 1 of 2 commercial nucleic acid amplification test assays, the Panther Fusion SARS-CoV-2 (Hologic Inc., <https://www.hologic.com>), as previously described (1), or the Panther Aptima SARS-CoV-2 (Hologic Inc.). The range of cycle threshold values for the rRT-PCR assays was <20 cycles to 45 cycles, and samples with a cycle threshold (C_t) value >40 were repeated and considered positive if reproducible. Results from the Panther Aptima assay were not included for analysis in the C_t comparison of this study because only a few patients were tested with this assay and the lack of robust correlation between C_t values and relative light units.

Two-Step Strand-Specific PCR

RNA extraction was performed from 400 μ L of respiratory specimen using the EZ1 instrument (QIAGEN, <https://www.qiagen.com>). Strand-specific PCR testing consisted of 2 sets of reactions. In the first set of reactions, reverse transcription with strand-specific primers converted SARS-CoV-2 RNA to complementary DNA (cDNA). A reverse envelope (*E*) gene primer generated cDNA to the plus strand (rtR), comprising both genomic RNA and messenger RNA. In a separate reaction, a forward *E* gene reaction generated cDNA to the minus strand (rtF), comprising minus genomic and subgenomic RNA. A third reverse transcription reaction without an added primer was used as a control for each nucleic acid eluate (background (no

primer) reverse transcription reaction; rtX). In the second step, the cDNA was amplified by real-time PCR in 3 separate reactions using the Rotor-Gene Q instrument (QIAGEN) with the Stanford *E* gene assay.

The oligonucleotide sequences of the primers and probes are listed in Appendix Table 1. Each reaction contained 12µL of 2X reaction mix and 1.2µL of enzyme mix, both from the Luna Universal Probe One-Step RT-qPCR kit (New England Biolabs, Inc., <https://www.neb.com>), 2 µL of reverse transcription primers (rtF, rtR or rtX), and 10µL eluate. Reverse transcription (RT) was performed on the Veriti instrument (Thermo Fisher Scientific, <https://www.thermofisher.com>) for 15 minutes at 60°C, followed by RT inactivation for 10 minutes at 80°C. This corresponds to a time-based RT separation, rather than an enzyme-based RT separation. This modification was performed to maximize the eluate input volume into the PCR reaction. The reaction tubes were then removed from the thermal cycler, and 2µL of the Stanford EUA *E* gene assay primer/probe mix was added in the same tubes, as previously described (2). In the second step, the amplification step was resumed in the same reaction tube by real-time PCR (rRT-PCR) using the Rotor-Gene Q instrument (QIAGEN). Thermal cycling involved 2 minutes at 94°C, followed by 45 cycles of 15 seconds at 94°C, 40 seconds at 55°C, and 20 seconds at 68°C. Each run included a SARS-CoV-2-negative control (pooled nasopharyngeal specimens confirmed negative by standard SARS-CoV-2 rRT-PCR), and 2 positive controls (in vitro transcribed plus- and minus-strand RNA). The minus-strand target was reported as detected if the rtF reaction showed an exponential curve with a $C_t < 45$ cycles, the rtX reaction was either not detected or detected ≥ 3 cycles later than the minus strand, and the rtR reaction was detected at a C_t less than that of the rtF reaction. Testing was performed as a laboratory-developed test for SARS-CoV-2 monitoring in a Clinical Laboratory Improvement Amendments–certified laboratory, and thus did not require emergency use authorization from the Food and Drug Administration.

Analytical Validation

To perform the analytical validation, minus- and plus-strand SARS-CoV-2 RNA was in vitro transcribed using the HiScribe T7 Quick Yield RNA Synthesis kit (New England Biolabs, Inc.) to be used as strand-specific control material. In vitro transcription was performed from T7-containing primers (Appendix Table 1), and one amplicon was generated for each direction. The

preparations were purified by DNase treatment followed by lithium chloride precipitation. The control RNA material was quantitated using the *E* gene assay with single-stranded DNA as standard. The control RNA material was run in triplicate from 6.0 to 1.0 log₁₀ copies/μL to evaluate linearity, and in triplicate at 10 and 100 copies/μL for 3 days to evaluate precision. The lower limit of detection (LLOD) was evaluated by testing 10 replicates each at 0.5, 1 and 2 copies/μL.

Research Cohort Selection (Retrospective Phase)

Patients whose initial strand-specific result was detected, and who had residual specimen leftover available from longitudinal samples, underwent repeat strand-specific testing. Repeat testing was not performed for individuals for whom initial strand-specific testing was negative. Specimens were stored at -80°C and thawed at the time of testing.

Clinical Cohort Selection (Prospective Phase)

Strand-specific rRT-PCR test results were reported in the electronic medical system. Results included the C_t value of the minus and plus strands, along with the respective interpretation as ‘detected’ or ‘not detected’ for each. Comments were autoappended to each strand-specific result to provide additional guidance on result interpretation. For positive results, the comment included that this was consistent with active virus replication. For negative results, the comment included a caution on the inability to rule out active virus replication on the basis of the strand-specific test result alone. Patients for whom testing had been ordered in error instead of routine RT-PCR for initial diagnosis were excluded. Test results were included for clinical decision-making on a case-by-case basis with support from the Stanford Infection and Prevention Control Program, and were not formally integrated into institutional guidelines at the time of this study. Chart review was performed as for the retrospective phase.

Statistical Analysis

The 95% LLOD was calculated by probit analysis using R version 4.0.2 (The R Project for Statistical Computing, <https://www.r-project.org>). Precision analysis was performed in Microsoft Excel (<https://www.microsoft.com>) as previously described (15). Clinical data statistical analysis was performed by χ^2 test or Fisher’s exact test when there were <5 datapoints per cell for categorical variables, or by the Mann-Whitney U test for continuous variables using

Stata version 15.1 (StataCorp LLC., <https://www.stata.com>). Multivariable logistic regression analysis was performed for the main clinical outcomes including age as a forced-in variable, and variables found to be significantly associated in bivariable analysis. Deming regression was performed to compare C_t results.

Results

Analytical Validation

The linear range extended from 1.0 to 6.0 \log_{10} copies/ μL for the minus and plus strands, with $R^2 = 0.99$ for both (Appendix Figure 1, panels A, B). Between-run and within-run precision showed coefficients of variation of $<2\%$ for both tested concentrations (Appendix Table 2). The 95% LLOD for the minus strand was calculated to be 0.5 RNA copies/ μL (5 copies/reaction; 10/10 samples detected). Background signal was observed for all minus strand in vitro transcribed control material, with an average signal difference of 11.8 cycles between the minus strand and background. The background signal corresponded to valid amplification curves, which could not be distinguished from true positives. Additional DNase treatment did not resolve this issue. Overall, these test performance characteristics were consistent with the performance seen with the Stanford *E* gene assay (B. Pinsky, unpub. data).

Clinical Results

The median C_t value from the standard nasopharyngeal rRT-PCR assay was 28.0 (interquartile range [IQR] 22.3–35.7) in the retrospective phase, and 33.4 (IQR 25.2–35.8) in the prospective phase, and the proportion of minus strand detection was significantly lower in the prospective phase (36.6% vs. 13.2%; $p < 0.01$). A subset of 13 persons in the retrospective phase and 9 persons in the prospective phase underwent repeat strand-specific rRT-PCR testing, of whom a total of 4 tested positive for minus-strand RNA, up to 5 days after the first test. In the retrospective phase, background signal was detected in 8 (8.6%) of 93 samples, with a median C_t value of 39.3 (IQR 38.7–39.9) (Appendix Table 3). In contrast, no background signal was observed in the prospective clinical cohort testing (Appendix Table 4).

In 2 patients in the prospective cohort, a negative result for minus-strand RNA facilitated the approval to proceed with procedures in the context of persistently positive SARS-CoV-2 rRT-PCR results. The first case included approval to proceed with surgery for a woman 18 years

of age for tumor investigation and management despite positive SARS-CoV-2 rRT-PCR results after >10 days of home isolation for coronavirus disease. The second case included facilitating approval to list an immunocompromised woman 56 years of age for organ transplant after >50 days of rRT-PCR positivity. Both of these presented with positive Aptima results, 1,135 and 1,027 relative light units, respectively.

References

- Hogan CA, Sahoo MK, Huang C, Garamani N, Stevens B, Zehnder J, et al. Comparison of the Panther Fusion and a laboratory-developed test targeting the envelope gene for detection of SARS-CoV-2. *J Clin Virol.* 2020;127:104383. PubMed <https://doi.org/10.1016/j.jcv.2020.104383>
- U.S. Food and Drug Administration. Stanford Health Care Clinical Virology Laboratory SARS-CoV-2 test EUA Summary. 2020 Nov 24 [cited 2020 Dec 2]. <https://www.fda.gov/media/136818/download>

Appendix Table 1. Strand-specific PCR primer and probe specifications in study on strand-specific reverse transcription PCR specific for detection of replicating severe acute respiratory syndrome coronavirus 2, California, USA, 2020*

| Name | Concentration, nmol/L | Sequence (5'→3') |
|-------------------------------------|-----------------------|--|
| Reverse transcription primers | | |
| nCoVsg_E.rtF | 400 | CGAACTTATGTA CTCTATTTCGTTTCGG |
| nCoVsg_E.rtR | 400 | AGAAGGTTTTACAAGACTCACGTT |
| <i>E</i> gene real-time PCR primers | | |
| SARS-CoV-2_E_Fwd | 400 | ACAGGTACGTTAATAGTTAATAGCGT |
| SARS-CoV-2_E_Rev | 400 | ATATTGCAGCAGTACGCACACA |
| RNase P Forward | 200 | AGATTTGGACCTGCGAGCG |
| RNase P Reverse | 200 | GAGCGGCTGTCTCCACAAGT |
| <i>E</i> gene real-time PCR probes | | |
| SARS-CoV-2_E_Prb-FAM | 200 | ACACTAGCCATCCTTACTGCGCTTCG |
| RNase P Cf-560 | 50 | TTCTGACCTGAAGGCTCTGCGCG |
| In vitro transcription primers | | |
| nCoVsg_E.FO.T7 | 400 | TAATACGACTCACTATAGGG CTTTGTAAGCACAAGCTGATGAGT |
| nCoVsg_E.RO | 400 | CCAGAAGATCAGGA ACTCTAGAAGA |
| nCoVsg_E.FO | 400 | CTTTGTAAGCACAAGCTGATGAGT |
| nCoVsg_E.Ro.T7 | 400 | TAATACGACTCACTATAGGG CCAGAAGATCAGGA ACTCTAGAAGA |

**E* gene, envelope gene; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2; Rtf, forward envelope gene reaction; RtR, reverse envelope gene reaction.

Appendix Table 2. Run results of study on strand-specific real-time reverse transcription PCR specific for detection of replicating severe acute respiratory syndrome coronavirus 2, California, USA, 2020*

| RNA | Mean cycle threshold value | Within run† | | Between runs† | | Total | |
|---------------------|----------------------------|-------------|-------|---------------|-------|-------|-------|
| | | SD | CV, % | SD | CV, % | SD | CV, % |
| Minus strand, cp/μL | | | | | | | |
| 100 | 30.7 | 0.15 | 0.49 | 0.52 | 1.70 | 0.54 | 1.75 |
| 10 | 34.1 | 0.30 | 0.87 | 0.53 | 1.54 | 0.58 | 1.70 |
| Plus strand, cp/μL | | | | | | | |
| 100 | 29.3 | 0.33 | 1.12 | 0.40 | 1.36 | 0.48 | 1.64 |
| 10 | 32.8 | 0.22 | 0.68 | 0.27 | 0.81 | 0.32 | 0.98 |

*cp/μL, copies per microliter; CV: coefficient of variation.

†The control RNA material was run in triplicate at 10 copies/μL and 100 copies/μL each day for 3 days to evaluate precision. Each run included a negative control (0cp/mL); the results of the negative control were not detected.

Appendix Table 3. Full results of the strand-specific real-time reverse transcription PCR for retrospective clinical samples from patients with severe acute respiratory syndrome coronavirus 2, California, USA, 2020*

| Patient ID | Sample type | Cycle threshold values | | | Cycle threshold by standard PCR† |
|------------|-------------|------------------------|--------------|-------------|----------------------------------|
| | | Background | Minus strand | Plus strand | |
| 1 | NP | ndet | 32.6 | 16.6 | 16.4 |
| 2 | NP | ndet | 37.4 | 23.5 | 26.4 |
| 3 | NP | ndet | 30.4 | 15.7 | 18.5 |
| 4 | NP | ndet | ndet | ndet | 37.5 |
| 5 | NP | ndet | ndet | 28.4 | 29.9 |
| 6 | NP | ndet | 37.8 | 24.9 | 25.8 |
| 7 | NP | ndet | 36.0 | 21.3 | 20.7 |
| 8 | NP | ndet | ndet | 31.7 | 32.5 |
| 9 | NP | ndet | 34.0 | 19.4 | 20.3 |
| 10 | NP | ndet | 39.2 | 22.2 | 21.3 |
| 11 | NP | 39.0 | ndet | 35.2 | 35.9 |
| 12 | NP | ndet | 34.9 | 18.9 | 15.6 |
| 13 | NP | ndet | 32.4 | 17.0 | 19.3 |
| 14 | NP | ndet | 34.1 | 17.5 | 15.3 |
| 15 | NP | ndet | ndet | 32.3 | 25.2 |
| 16 | NP | ndet | ndet | 30.8 | 35.9 |
| 17 | NP | ndet | 32.6 | 16.6 | 15.0 |
| 18 | NP | ndet | ndet | 30.5 | 33.2 |
| 19 | NP | ndet | 39.1 | 25.9 | 24.8 |
| 20 | NP | 39.7 | ndet | 26.0 | 23.2 |
| 21 | NP | ndet | 35.9 | 26.6 | 27.1 |
| 22 | NP | ndet | ndet | 28.1 | 27.0 |
| 23 | NP | ndet | 30.8 | 20.2 | 18.9 |
| 24 | NP | ndet | ndet | 39.3 | 38.6 |
| 25 | NP | ndet | ndet | 29.3 | 29.8 |
| 26 | NP | ndet | ndet | ndet | 34.5 |
| 27 | NP | ndet | ndet | ndet | 40.5 |
| 28 | NP | ndet | ndet | 34.7 | 32.6 |
| 29 | NP | ndet | 34.7 | 20.8 | 23.5 |
| 30 | NP | ndet | ndet | 27.1 | 27.3 |
| 31 | NP | ndet | ndet | 28.7 | 30.1 |
| 32 | NP | ndet | ndet | 34.2 | 35.4 |
| 33 | NP | 37.5 | 27.4 | 14.3 | 16.4 |
| 34 | NP | ndet | ndet | 34.5 | 36.2 |
| 35 | NP | ndet | ndet | 38.7 | 39.5 |
| 36 | NP | ndet | ndet | 32.6 | 35.7 |
| 37 | NP | ndet | ndet | 32.9 | 36.8 |
| 38 | NP | ndet | ndet | 33.0 | 36.7 |
| 39 | NP | ndet | ndet | 36.3 | 41.8 |
| 40 | NP | ndet | ndet | 35.5 | 39.5 |
| 41 | NP | ndet | ndet | 27.4 | 30.0 |
| 42 | NP | ndet | 33.2 | 18.5 | 19.0 |
| 43 | NP | ndet | ndet | 31.9 | 31.6 |
| 44 | NP | ndet | 31.2 | 20.3 | 23.3 |
| 45 | NP | ndet | 38.6 | 24.4 | 24 |
| 46 | NP | ndet | ndet | 22.2 | 23.9 |
| 47 | NP | ndet | ndet | 36.7 | 36.5 |

| | | | | | |
|----|-------|------|------|------|------|
| 48 | NP | 38.8 | ndet | 28.5 | 29.2 |
| 49 | NP | ndet | ndet | 35.5 | 33.4 |
| 50 | NP | ndet | ndet | 32.1 | 35.7 |
| 51 | NP | ndet | ndet | 32.2 | 32.2 |
| 52 | NP | 38.5 | 31.9 | 15.9 | 16.3 |
| 53 | NP | ndet | ndet | 35.5 | 33.2 |
| 54 | NP | 41.1 | ndet | 31.6 | 32.3 |
| 55 | NP | ndet | 37.5 | 22.1 | 23.9 |
| 56 | NP | ndet | ndet | 34.5 | 34.1 |
| 57 | NP | 39.9 | 32.2 | 16.8 | 13.5 |
| 58 | NP | ndet | ndet | 28.9 | 25.1 |
| 59 | OP | ndet | ndet | 28.4 | 24.4 |
| 60 | NP | ndet | ndet | 28.4 | 26.5 |
| 61 | NP | ndet | 32.2 | 19.1 | 16.7 |
| 62 | NP | ndet | 36.7 | 23.6 | 24.9 |
| 63 | NP | ndet | ndet | 29.1 | 28.0 |
| 64 | NP | ndet | ndet | 27.1 | 24.9 |
| 65 | Nasal | ndet | 39.3 | 24.0 | 19.3 |
| 66 | OP | ndet | ndet | 29.5 | 23.2 |
| 67 | OP | ndet | 38.8 | 22.5 | 18.9 |
| 68 | NP | ndet | 36.4 | 21.2 | 20.8 |
| 69 | NP | ndet | 39.9 | 21.5 | 21.2 |
| 70 | NP | ndet | 37.4 | 23.3 | 22.3 |
| 71 | NP | ndet | ndet | 28.7 | 28 |
| 72 | Nasal | ndet | 39.8 | 28.6 | 24.1 |
| 73 | NP | ndet | ndet | 32.6 | 33.1 |
| 74 | NP | ndet | ndet | 36.3 | 36.9 |
| 75 | NP | ndet | ndet | 28.6 | 29.4 |
| 76 | NP | ndet | 39.3 | 24.8 | 22.2 |
| 77 | NP | ndet | ndet | 36.5 | 36.4 |
| 78 | NP | ndet | ndet | 38.0 | 42.7 |
| 79 | NP | ndet | ndet | 35.3 | 36.4 |
| 80 | NP | ndet | ndet | ndet | 37.1 |
| 81 | NP | ndet | 36.1 | 18.0 | 18.3 |
| 82 | OP | ndet | 39.4 | 26.3 | 21.8 |
| 83 | NP | 40.1 | 33.3 | 14.4 | 16.8 |
| 84 | NP | ndet | ndet | 29.6 | 30.5 |
| 85 | NP | ndet | ndet | 24.6 | 25.0 |
| 86 | OP | ndet | ndet | 34.3 | 27.7 |
| 87 | NP | ndet | ndet | 36.7 | 35.5 |
| 88 | NP | ndet | ndet | ndet | 35.9 |
| 89 | NP | ndet | ndet | ndet | 37.3 |
| 90 | NP | ndet | ndet | ndet | 38.5 |
| 91 | OP | ndet | ndet | ndet | 35.9 |
| 92 | NP | ndet | ndet | 34.8 | 30.9 |
| 93 | OP | ndet | ndet | ndet | 37.5 |


*ndet, not detected; NP, nasopharyngeal; OP, oropharyngeal.

Appendix Table 4. Full results of the strand-specific real-time reverse transcription PCR assay for prospective clinical samples from patients with severe acute respiratory syndrome coronavirus 2, California, USA, 2020*


| Patient ID | Sample type | Cycle threshold values | | | Cycle threshold by standard PCR or relative light units by transcription-mediated amplification† |
|------------|-------------|------------------------|--------------|-------------|--|
| | | Background | Minus strand | Plus strand | |
| 94 | NP | ndet | ndet | 39.6 | ndet |
| 95 | NP | ndet | ndet | 35.9 | 34.2 |
| 96 | NP | ndet | ndet | 34.6 | 1,135† |
| 97 | NP | ndet | ndet | ndet | 40.1 |
| 98 | NP | ndet | ndet | ndet | 584† |
| 99 | NP | ndet | ndet | 31.1 | NA |
| 100 | NP | ndet | ndet | 31.2 | 31.0 |
| 101 | NP | ndet | ndet | ndet | NA |
| 102 | NP | ndet | 28.7 | 16.1 | NA |
| 103 | NP | ndet | ndet | 38.7 | 37.3 |
| 104 | NP | ndet | ndet | ndet | NA |
| 105 | NP | ndet | ndet | 32.2 | 1,144† |
| 106 | NP | ndet | ndet | 38.9 | 35.2 |
| 107 | NP | ndet | ndet | 35.2 | 1,057† |
| 108 | NP | ndet | ndet | 34.9 | 21.4 |
| 109 | NP | ndet | ndet | 34.5 | 34.4 |
| 110 | Nasal | ndet | ndet | 36.5 | NA |
| 111 | NP | ndet | 27.9 | 13.0 | NA |
| 112 | NP | ndet | ndet | 37.6 | 36.6 |
| 113 | Nasal | ndet | ndet | ndet | 38.0 |
| 114 | NP | ndet | ndet | 38.4 | 35.8 |
| 115 | NP | ndet | ndet | 25.9 | 27.2 |
| 116 | NP | ndet | ndet | 34.8 | 33.6 |
| 117 | NP | ndet | 39.0 | 23.7 | 18.5 |
| 118 | NP | ndet | 33.2 | 19.2 | 20.8 |
| 119 | NP | ndet | ndet | ndet | 1,140† |
| 120 | NP | ndet | ndet | 26.4 | 25.2 |
| 121 | NP | ndet | ndet | ndet | 1,027† |
| 122 | NP | ndet | ndet | 34.7 | 35.1 |
| 123 | NP | ndet | ndet | 35.3 | 34.7 |
| 124 | NP | ndet | ndet | 38.5 | 783† |
| 125 | NP | ndet | ndet | 34.0 | 776† |
| 126 | NP | ndet | ndet | 30.9 | 33.4 |
| 127 | NP | ndet | 38.6 | 31.9 | 34.5 |
| 128 | NP | ndet | ndet | 37.3 | NA |
| 129 | NP | ndet | 34.3 | 21.1 | 22.6 |
| 130 | NP | ndet | ndet | 34.7 | 32.1 |
| 131 | Nasal | ndet | ndet | 27.0 | 29.3 |
| 132 | NP | ndet | ndet | ndet | 1,282† |
| 133 | NP | ndet | ndet | ndet | 37.6 |
| 134 | NP | ndet | ndet | 26.2 | 28 |
| 135 | NP | ndet | ndet | ndet | 18.6 |
| 136 | NP | ndet | ndet | 37.5 | 35.8 |
| 137 | NP | ndet | ndet | 32.77 | 33.1 |
| 138 | NP | ndet | ndet | ndet | 36.9 |
| 139 | NP | ndet | ndet | 35.48 | 1,084† |
| 140 | NP | ndet | ndet | 34.94 | 36.8 |
| 141 | NP | ndet | 27.61 | 16.09 | 18.1 |
| 142 | NP | ndet | ndet | 35.22 | 36.1 |
| 143 | NP | ndet | ndet | ndet | 38.4 |
| 144 | NP | ndet | ndet | ndet | 957† |
| 145 | NP | ndet | ndet | 34.04 | 30.94 |
| 146 | NP | ndet | ndet | 36.55 | NA |

*NA, not available; ndet, not detected; NP, nasopharyngeal.

†Samples were tested by real-time reverse transcription PCR or by transcription-mediated assay at the Stanford Clinical Virology Laboratory, Stanford, CA, USA. Crosses (†) indicate samples tested on the transcription-mediated assay, for which the results are expressed as relative light units. NA values indicate tests that were performed at external laboratories.



Appendix Figure 1. Linearity of Ct values obtained by strand-specific real-time reverse transcription PCR for the SARS-CoV-2 (A; $y = -3.34x + 34.62$; $R^2 = 0.99$) minus strand and (B; $y = -3.32x + 34.80$; $R^2 = 0.99$) plus strand, California, USA, 2020. Ct, cycle threshold; SARS-CoV-2, severe acute respiratory syndrome coronavirus 2.



Appendix Figure 2. Ratio of minus:plus strand viral load in samples from patients with severe acute respiratory syndrome coronavirus 2 infection, California, USA, 2020. The dataset included 38 samples for which time to symptom onset data were available. Three outliers were excluded from this figure for improved visualization: 1 sample collected ≤ 7 days after symptom onset ($y = 3.44$), and 2 samples collected 8–14 days after symptoms onset ($y = 1.18$ and $y = 0.21$).